



INSTRUCTIONS FOR USE

YA23L-D Steel Security Safe **Weatherproof version** **PULL OUT DEPOSIT DRAWER**

When any person is responsible for the management and / or disposal of contaminated waste, it is recommended they comply with Occupational Health & Safety Guidelines for the safe handling of Sharps. Suggested OH & S Procedures are attached to these Instructions .

Your YA23L-D Security Safe is supplied with an IDC 23 litre Sharps Container Base Unit (code RE20LR) already installed in the Steel Safe. Contaminated waste (needles / syringes, syringe packs) are safely disposed of by depositing directly into the PULL OUT DRAWER / CHUTE located at the top front of the steel Security Safe.

FITTING A SHARPS CONTAINER (Base Unit) INTO THE STEEL SAFE:

- Please read the assembly sheet (yellow) for the RE20LR Sharps Container.
- The front of the Sharps Container is designated by the bio-hazard label. For use with the YA23L-D Security Safe the RE20LR sharps container is supplied with a solid lid
- The Steel safe door opens with the key provided .
- Place the Sharps Container Base Unit inside the Safe with the handle (red) positioned over the front. The handle in this position makes it safer & easier to remove the Base Unit from the Steel Safe. If correctly installed, the centre open top of the Sharps Container will be aligned directly underneath the deposit chute.

If the Sharps Container Base Unit is installed correctly, needles / syringes & syringe packs will deposit successfully into the Sharps Container Base Unit each time.

- The Steel Safe is now ready to use and must be locked.

WHEN THE SHARPS CONTAINER (Base Unit) IS FULL:

- Unlock Safe with key.
- Make sure the Sharps Container Base Unit is safe to remove.
- After removal from the Steel Safe, place the solid lid on top of the sharps container Base Unit
- Permanently seal the Sharps Container by FIRMLY pressing the Solid Lid onto the Base Unit of the Sharps Container.
- The Sharps Container is now LOCKED and ready for safe transport to the place of final disposal.
- An empty replacement Sharps Container Base Unit can now be installed in the Security Safe.

For your personal protection, and the safety of others;

Please remember that: Polymer products, although puncture resistant, are not entirely puncture proof. Always follow the guidelines set down by the Infection Control Association (or equivalent). A sharps container should never be filled above the indicator fill line. An article must Never be forced into a Sharps Container. Syringes must Never, Ever, be deposited into a Sharps Container with the needle pointing upwards.

IDEAS DEVELOPMENT CORPORATION PTY. LIMITED

ABN 84 078 760 504 ACN 078 760 504

Incorporating IDC Medical

Phone : 02 4321 0299 ~ Fax : 02 4321 0288

Web site : www.idcmedical.com

e-mail : sharps@idcmedical.com

OFFICE & WAREHOUSE

MAILING ADDRESS

16/13 Gibbens Road, West Gosford NSW

PO Box 6290, Kincumber NSW 2251